

Revision History

Date	Version	Revision
2020/1/12	5.09.01.07	1. Fixed a measure bug
2018/5/11	5.09.01.05	1. The probe attenuation factor is considered in the trend diagram mode. 2. Digital filtering state remains even in scope stop state. 3. The switchboard can record the configuration of the math waveform. 4. The Math waveform takes into account the probe coefficient configuration. 5. Improved Win10/EasyScopeX connectivity. 6. Improved vertical cursor measurements when changing scales on fixed/stopped waveforms 7. Enabled self-calibration under Normal or Single mode. 8. Improved edge trigger performance for single and normal mode 9. Increased Single trigger data depth in long storage mode(from 20kto2Mpts). 10. Increased stability of trend map transfer of CSV data to USB memory devices > 4 GB 11. Optimize the display algorithm of multi-meter trend chart and oscilloscope trend chart. Solves the slow response of keys after long recording time. 12. Enhanced FFT data usage. Long storage features 40 times the corresponding frequency points used in ordinary storage. 13. Improved read/write timeout to the U disk. 14. Time stamp information added to the saved CSV file.
2016/01/21	3.01.02.02R7	1. A formal version of the previous version

Compatibility between Versions

Source Version	Object Version	Compatibility
<5.09.01.07	5.09.01.07	Yes

Update instructions

◆ **Very important!**

If your firmware is not 5.09.01.05, Please upgrade it to 5.09.01.05 first