

Oscilloscope Specifications: What does “div” mean in the specifications?

February 27, 2019

The term “div” for most SIGLENT specifications is equal to the Vertical Scale value (Commonly shown in units of Volts/div).

Take the specification for the DC Gain Accuracy of a SIGLENT SDS1052DL+

For ranges ≤ 100 mV/div, the DC Gain Accuracy is given as

$$\pm [3\% \times (|\text{reading}| + |\text{offset}|) + 1\% \times |\text{offset}| + 0.2 \times \text{div} + 2 \text{ mV}]$$

If:

$$\text{Reading} = 0.5 \text{ V}$$

$$\text{Offset is } 0.1 \text{ V}$$

$$\text{Vertical scale (div)} = 100 \text{ mV/div} = 0.1 \text{ V/div,}$$

Then DC Gain is calculated as follows:

$$\pm [3\% \times (|\text{reading}| + |\text{offset}|) + 1\% \times |\text{offset}| + 0.2 \times \text{div} + 2 \text{ mV}] =$$

$$\pm [3\% \times (|0.5 \text{ V}| + |0.1 \text{ V}|) + 1\% \times |0.1 \text{ V}| + 0.2 \text{ div} \times (0.1 \text{ V/div}) + 0.002 \text{ V}] =$$

$$\pm [3\% \times (0.6 \text{ V}) + 0.0001 \text{ V} + 0.02 \text{ V} + 0.002 \text{ V}] = 0.018 \text{ V} + 0.0001 \text{ V} + 0.02 \text{ V} + 0.002 \text{ V} = 0.0401 \text{ V}$$



North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139

Tel: 440-398-5800

Toll Free: 877-515-5551

Fax: 440-399-1211

info@siglent.com

www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH

Staezlinger Str. 70

86165 Augsburg, Germany

Tel: +49(0)-821-666 0 111 0

Fax: +49(0)-821-666 0 111 22

info-eu@siglent.com

www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.

Blog No.4 & No.5, Antongda Industrial Zone,

3rd Liuxian Road, Bao'an District,

Shenzhen, 518101, China.

Tel: + 86 755 3661 5186

Fax: + 86 755 3359 1582

sales@siglent.com

www.siglent.com/ens