

SDS1000X-E New Fixed Position Zoom

February 08, 2018

This feature allows the user to adjust the position selections for horizontal and vertical zooming.

Traditionally, most scopes allow you to zoom around a fixed position. This is typically the center of the display (which is the standard operation on all SIGLENT scopes).


We refer to this as fixed delay for horizontal and fixed offset for vertical zooming. In this configuration, the delay and offset values remain fixed which causes any vertical or horizontal scale changes to be centered on the middle of the display.

Here is a screen image of an SDS1202X-E with an input signal:




Note the location of the horizontal and vertical position badges.

With the traditionally fixed offset zoom, when you adjust the vertical and horizontal scales, you can see that the waveform expands, but it does so around the center of the display. The vertical and horizontal positions change:



When we zoom around fixed positions, the horizontal offset and vertical positions are fixed on the display and the waveform is expanded around those points as shown here:



This new feature allows you to select a fixed reference position for both the vertical and horizontal offset positions for the zoom. This makes for faster troubleshooting by minimizing the number of adjustments required to zoom on the area of interest.



North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free:877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel:+ 86 755 3661 5186
Fax:+ 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens