

SDS6000A Typical Noise Floor

June 24, 2022

Here are the typically expected noise floor numbers for the SDS6000A Oscilloscope:

Vertical Scale Noise Floor (rms@50 ohms)

0.5 mV/div	135 uV
1 mV/div	135 uV
2 mV/div	136 uV
5 mV/div	200 uV
10 mV/div	325 uV
20 mV/div	490 uV
50 mV/div	1.10 mV
100 mV/div	2.01 mV
200 mV/div	5.10 mV
500 mV/div	11.60 mV
1 V/div	20.95 mV

Note: Uses StdDev measurement @ 5 GSa/s, 10 MPts/ch



North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free:877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel:+ 86 755 3661 5186
Fax:+ 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens