

Typical SDG6000X Phase Noise

November 10, 2021

Phase Noise

Phase noise is the random fluctuations of the phase of a signal in the frequency domain. Jitter is a similar measurement in the time domain. They are both caused by the random thermal fluctuations of the oscillator used as a timing reference in measurement instruments, signal sources, and other circuits that rely on oscillators or repetitive signals for timing and clocking functions.

These fluctuations cause the phase of the output clock signal to vary with time, very similar to jitter in a time-based system. This causes a widening of the signal when viewed in the frequency domain (see Figure 1 below). This can cover up any small signals that may be near the frequency of interest when measuring or can mask or cover required signals if the source phase noise is too high. For meaningful measurements or sources, select an instrument with lower phase noise than the signal source or measurement instrument that you are testing.






FIGURE 1

Below are some typical phase noise measurements of the SIGLENT SDG6000X Series of Arbitrary Waveform Generators.

Typical SDG6000X Function Generator Phase Noise @ 10 MHz carrier:



Typical SDG6000X Function Generator Phase Noise @ 100 MHz carrier:



Typical SDG6000X Function Generator Phase Noise @ 200 MHz carrier:





North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free: 877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel: + 86 755 3661 5186
Fax: + 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens